


7-01-001

MEDICATION ADMINISTRATION VIA FREEDOM 60 PUMP



SUPPLIES:

- Syringe of medication
- Freedom 60 tubing (tubing to be changed DAILY)
- (2) Sodium Chloride (Saline) syringes
- (1) Heparin syringe (if necessary)
- (1) Blue cap

PROCEDURE:

1. Gather equipment. Clean work area. Wash hands.
2. Prepare your flushes as instructed by your nurse.
3. **Verify that tubing is labeled "Freedom 60 tubing."** Remove tubing from package.
4. Attach tubing to end of syringe.
5. Freedom pump should be rewound for this dose of medication, which will have moved the black tab to the end of its track.
6. Insert syringe into the clear cover on the pump. Insert plunger end first.
7. Clean the injection cap on the end of your catheter with a new alcohol wipe vigorously for 30 seconds and let air dry for 60 seconds.
8. Flush your catheter with Sodium Chloride as instructed by your nurse.



7-01-001

MEDICATION ADMINISTRATION VIA FREEDOM 60 PUMP

9. To prime the tubing, turn on the pump until a drop of fluid is noted at the end of the tubing.
10. Clean the injection cap with an alcohol wipe vigorously for 30 seconds and let air dry for 60 seconds, then attach the Freedom 60 tubing to the end of your injection cap.

At the end of the infusion:

1. When the syringe is empty, the infusion is finished. Turn pump to "0" (off).
2. Disconnect the tubing from your injection cap.
3. Keeping the end of the tubing sterile, place a blue cap on the end of the tubing (if another dose is to be administered the same day).

NOTE: The Freedom 60 tubing is to be changed DAILY. You will reuse the tubing when administering more than one infusion per day.

4. Rewind the knob clockwise, moving the black tab to the end of the track.
5. Remove the syringe from the pump. Leave empty syringe attached to the tubing.
6. Clean the injection cap on the end of your catheter with an alcohol wipe vigorously for 30 seconds and let air dry for 60 seconds.
7. Flush your catheter with Sodium Chloride as instructed by your nurse.

IF NEEDED and instructed by your nurse:

8. Clean the injection cap on the end of your catheter with a new alcohol wipe vigorously for 30 seconds and let air dry for 60 seconds. Flush your catheter with Heparin as instructed by your nurse.
9. When the next dose is due, remove used syringe from tubing, and replace with a new syringe of medicine. Repeat process of administration of medication.

SASH FLUSH PROCEDURE:

- S** Sodium Chloride ___ ml
- A** Administer medication as instructed by your nurse
- S** Sodium Chloride ___ ml
- H** Heparin ___ ml (if required)

Supplies and waste can be double bagged and disposed of with your household garbage.

IMPORTANT SAFETY ALERT:

TO AVOID RAPID INFUSION OF YOUR MEDICATION, OR AN ERROR, PLEASE VERIFY THE TUBING IS LABELED "FREEDOM 60" BEFORE INFUSING YOUR MEDICATION.

Revised 8/2015, 11/2016

Home Infusion & Specialty Pharmacy
CHARTWELLSM